



# Certificate of Accreditation: Supplement

**STK TECHNOLOGY CO., LTD.**  
**Production Department Test House**  
**2443-1 Misa Oita-shi, Oita 870-0108**  
**Kazuhisa Eto Phone: 097-521-1437**

*Accreditation is granted to the facility to perform the following testing:*

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND GOVERNMENT ORDINANCE STANDARD VALUE
Electrical <sup>F</sup>	Semiconductor device	Temperature setting in the chamber of Burn-In Machine	Procedures for High Temperature Operating Life Test (SBS21019) STT-01-RT-001  On the basis of: JESD22-A108-4.2.3.2 Standard  Thermograph	125 °C ± 3 °C
		Voltage setting of Burn-In Machine	Multi-meter R6452A	1.6 V ± 0.3 V 3.7 V ± 0.3 V

- The presence of a superscript F means that the laboratory performs testing of the indicated parameter at its fixed location. Example: Outside Micrometer <sup>F</sup> would mean that the laboratory performs this testing at its fixed location.